



No.C-14011/CPCB/PT-30/Water/65

Date: 30th April, 2014

LABORATORY DIVISION – WATER LABORATORY
30th PROFICIENCY TESTING PROGRAMME – METAL – DECEMBER, 2013

REPORT

Laboratory Address : The Laboratory Incharge
Maharashtra Pollution Control Board,
Jog Centre, 3rd Floor, Old Pune - Mumbai Road
Wakdewdi
PUNE - 411 003, MAHARASHTRA

Laboratory Code : 65

S.No.	Test Parameter	Unit	Sample	Reference Value	SD (NIQR Basis)	Lab's result	Lower Accept. Limit	Higher Accept. Limit	Robust Z-score	Remarks (Y/N/NR)
1.	IRON	mg/l	A	1.11	0.13	1.09	0.86	1.36	-0.15	Y
			B	3.34	0.42	3.29	2.50	4.18	-0.08	Y
2.	CADMIUM	mg/l	A	0.34	0.04	0.37	0.27	0.41	0.68	Y
			B	0.93	0.06	0.98	0.80	1.06	0.98	Y
3.	COPPER	mg/l	A	1.07	0.09	1.06	0.88	1.26	0.35	Y
			B	4.32	0.73	4.34	3.85	4.79	0.06	Y
4.	NICKEL	mg/l	A	1.13	0.09	1.05	0.94	1.31	-0.51	Y
			B	3.34	0.26	3.16	2.81	3.87	-0.68	Y
5.	LEAD	mg/l	A	1.67	0.22	1.50	1.24	2.10	-0.79	Y
			B	3.34	0.31	2.98	2.72	3.96	-1.16	Y
6.	MAGNESE (Mn) mg/l	mg/l	A	0.54	0.06	0.58	0.43	0.65	0.68	Y
			B	3.30	0.23	3.27	2.84	3.75	-0.13	Y
Overall performance in %										100.0

Note: Acceptable limits are fixed based on **Reference value** (Median) \pm 2 **Standard Deviation** (NIQR basis) i.e. within \pm 2 **Z Scores** values for both sample A & B individually.

IQR (Inter Quartile Range) = Q3-Q1; **Q1** = 1st Quartile; **Q3** = 3rd Quartile
NIQR (Normalized IQR) = IQRx0.7413 = 1 Standard Deviation
Z Score = (Observed value - Reference Value) / 1 Standard Deviation
Y - Within Acceptable Limit; **N** - Not within acceptable limit; **NR** - Not Reported

(Dr. P. K. Behera)
Scientist 'C' &
I/c. Water Lab