

MINUTES OF TENDER COMMITTEE MEETING
(e-Tender no. 1008; "Selection of Supplier for Supply and providing O & M Support for Sound Level Meters to MPCB in the State of Maharashtra")

Date & Time : 5th October 2019 at 0230 Hrs.

Venue : Conference Room, MPC Board, Sion.

Committee Members:

MPCB : JD (APC), CAO, LO, SA
CPCB : Dr. S K Sharma
NEERI : Dr. K V George, Head APC Division CSIR-NEERI
MAHAGENCO: Mr. P.P. Bhoje

Agenda : Pre-BID meeting for the e-tender: "Selection of Supplier for Supply and providing O & M Support for Sound Level Meters to MPCB in the State of Maharashtra".
(E-tender no: 1008; Tender reference no: MPCB / JD (APC) /05/2019-20)

The pre-BID meeting was held to reply the queries raised as per tender clause: 6.1 (Enquiries and Clarifications) by the different proposed bidders. Following Bidders were present in the pre-BID meeting.

- 1) M/s.Nevco Engineers Pvt. Ltd.;
- 2) M/s. Saksham Scientific Instruments Pvt. Ltd;

The Committee clarified the queries raised by the bidders and same are recorded. The Queries raised by **M/s Nevco Engineers Pvt. Ltd** and **M/s Saksham scientific Instruments Ltd** with tender committee remarks are placed at **Annexure-I**.

The committee has also decided that, pre-BID minutes of the meeting should be treated as part of the Tender Conditions of e-tender 1008 and shall be binding to both, MPCB and Bidders.

The committee has also decided that, the queries raised / received after this pre-BID meeting, shall not be considered for e-tender bidding.

The meeting ended with vote of thanks to the chair.

[Handwritten Signature]
5/10/19

[Handwritten Signature]
9/10/19

Annexure-I

• Queries raised by **M/s Nevco Engineers Pvt. Ltd:**

| Sr No. | E-Tender Page and Clause No: | Description in E-Tender | Clarification Sought | Remark |
|--------|------------------------------|--|--|---|
| 1. | - | - | Details of Phase-out equipment's along with contact numbers so that we can visit and check | <ul style="list-style-type: none"> The Board is having Type I/II (Make-Quest, Model-QC-1D- Purchased in 2007-08) equipment's to be given under Buy-Back scheme. Buy back phased out equipment's on "as is where is basis". Bidder should visit MPCB offices in Maharashtra and get the list of equipment's to be phased out under Buy-Back scheme. Bidder has to quote rates for buy-back offered per unit basis. |
| 2. | 43-Annexure-2 | Total Measurement Range 20 to 140 dB RMS (Single Range) GPS Facility | 25-140 dB Range | No change in the tender |
| 3. | 43-Annexure-2 | Inbuilt GPS in Sound Level Meter, GPS location to get printed on report generated in Field | We don't have inbuilt GPS in Sound Level Meter | No change in the tender |
| 4. | 44-Annexure-2 | User settable timer duration -5 sets of users selectable on/off timer and repeat function | 2 sets | No change in the tender |

Signature
5/10/19

Signature
9/11/19

• Queries raised by **M/s Saksham scientific Instruments Ltd:**

| Sr No. | E-Tender Page and Clause No: | Description in E-Tender | Clarification Sought | Remark |
|--------|------------------------------|-------------------------------|--|--|
| 1. | Page no 29 Clause no 22 | Annual Maintenance Contract | As per clause no. 22 AMC period is 2 years, while as per Annexure-4, AMC period is 5 Years | Kindly read as Warranty Support of Sound Level Monitoring Kits for TWO (2) years and AMC for THREE (3) years, There is no change in Tender conditions |
| 2. | Page no 29 Clause no 22 | Annual Maintenance Contract | clause no. 22 AMC period is 2 years, while as per Annexure-4.2, AMC period is 3 Years | Kindly read as Warranty Support of Sound Level Monitoring Kits for TWO (2) years and AMC for THREE (3) years. |
| 3. | Page no 08 Clause no 5.7 | Schedule of Activities | clause no. 5.7, last date for submission of bid is 09/10/2019, which gives only one working day after pre-bid meeting which has hold on 05/10/2019 | There will be NO change in schedule of activities For any change in schedule, Corrigendum's from time to time, please visit https://maharashtra.etenders.in . |
| 4. | Other Corrections | Annexure-4 (Price Bid Format) | 5 Years AMC/ CAMC Price | 5 Years AMC/ CAMC Price- Should be read as AMC-3 Years Note: 3 Years Warranty and 3 years AMC |

[Handwritten signature]
5/10/19

[Handwritten signature]
05/10/19